

**Deutsche WindGuard
Wind Tunnel Services GmbH, Varel**



accredited by the / akkreditiert durch die

Deutsche Akkreditierungsstelle GmbH

as calibration laboratory in the / als Kalibrierlaboratorium im

Deutschen Kalibrierdienst



Deutsche
Akkreditierungsstelle
D-K-15140-01-00

1522485
D-K-
15140-01-00
07/2015

Calibration certificate
Kalibrierschein

Calibration mark
Kalibrierzeichen

Object <i>Gegenstand</i>	3D Sonic Anemometer
Manufacturer <i>Hersteller</i>	Gill Instruments UK-Hampshire S041 9EG
Type <i>Typ</i>	1590-PK-020
Serial number <i>Fabrikat/Serien-Nr.</i>	Y152201
Customer <i>Auftraggeber</i>	Gill Instruments UK-Hampshire S041 9EG
Order No. <i>Auftragsnummer</i>	M42664
Project No. <i>Projektnummer</i>	VT150742
Number of pages <i>Anzahl der Seiten</i>	5
Date of Calibration <i>Datum der Kalibrierung</i>	31.07.2015

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI).

Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.

This calibration certificate may not be reproduced other than in full except with the permission of both the German Accreditation Body and the issuing laboratory. Calibration certificates without signature are not valid. This calibration certificate has been generated electronically.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Deutschen Akkreditierungsstelle als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift haben keine Gültigkeit. Dieser Kalibrierschein wurde elektronisch erzeugt.

Date
Datum

17.08.2015

Head of the calibration laboratory
Leiter des Kalibrierlaboratoriums

Dipl. Phys. Dieter Westermann

Person in charge
Bearbeiter

Dipl.-Ing. (FH) Catharina Herold

1522485
D-K-
15140-01-00
07/2015

Calibration object
Kalibiergegenstand

3D Sonic Anemometer

Calibration procedure
Kalibrierverfahren

- Deutsche WindGuard Wind Tunnel Services: Calibration of anemometers; Version 1.0 (2014)
- Based on following standards:
- MEASNET: Anemometer calibration procedure
 - IEC 61400-12-1: Power performance measurements of electricity producing wind turbines
 - IEC 61400-12-2: Power performance of electricity producing wind turbines based on nacelle anemometry
 - ISO 3966: Measurement of fluid in closed conduits
 - ISO 16622: Meteorology - Sonic anemometers/thermometers

Place of calibration
Ort der Kalibrierung

Windtunnel of Deutsche WindGuard WindTunnel Servies GmbH, Varel

Test conditions
Messbedingungen

wind tunnel area	10000 cm ²
anemometer frontal area	220 cm ²
diameter of mounting pipe	50 mm
blockage ratio ¹⁾	0.022 [-]
software version	7.64

¹⁾ Due to the special construction of the test section no blockage correction is necessary.

Ambient conditions
Umgebungsbedingungen

air temperature	23.2 °C ± 0.1 °C
air pressure	1019.4 hPa ± 0.3 hPa
relative air humidity	57.6 % ± 2.0 %

Measurement uncertainty
Messunsicherheit

The expanded uncertainty assigned to the measurement results is obtained by multiplying the standard uncertainty by the coverage factor k = 2. It has been determined in accordance with DAkkS-DKD-3. The value of the measurand lies within the assigned range of values with a probability of 95%.
The reference flow speed measurement is traceable to the German NMI (Physikalisch-Technische Bundesanstalt) standard for flow speed. It is realized by using a PTB owned and calibrated Laser Doppler Anemometer (Standard Uncertainty 0.2 %, k=2)

Latest accreditation
Letzte Akkreditierung

04/2014

Additional remarks
Zusätzliche Anmerkungen

Orientation: 180°

1522485
D-K-
15140-01-00
07/2015

Calibration result
Kalibrierergebnis

Sensor v_hor	Sensor dir	Sensor v_vert	Tunnel speed	Uncertainty (k=2)
m/s	deg	m/s	m/s	m/s
3.948	180.787	-0.010	3.953	0.050
5.876	180.169	-0.022	5.893	0.050
7.869	180.363	-0.026	7.890	0.050
9.809	180.506	-0.025	9.842	0.050
11.796	180.825	-0.023	11.848	0.050
13.701	180.831	-0.015	13.769	0.050
15.650	181.000	-0.002	15.739	0.050
14.714	180.963	-0.006	14.803	0.050
12.749	180.975	-0.022	12.817	0.050
10.761	180.688	-0.026	10.813	0.050
8.840	180.525	-0.025	8.878	0.050
6.889	180.131	-0.024	6.919	0.050
4.880	180.381	-0.017	4.903	0.050

File: 1522485

Linear regression analysis	Slope	1.00687 (m/s)/(m/s) ± 0.00056 (m/s)/(m/s)
	Offset	-0.0224 m/s ± 0.006 m/s
	Standard error (Y)	0.008 m/s
	Correlation coefficient	0.999998

Remarks
The calibrated sensor complies with the demanded linearity of MEASNET



1522485
D-K-
15140-01-00
07/2015

Graphical representation of the result
Grafische Darstellung des Ergebnisses




Photo of the measurement setup
Foto des Messaufbaus



Remark: The proportions of the set-up may not be true to scale due to imaging geometry.

1522485
D-K-
15140-01-00
07/2015

Sensor config during calibration
Sensorkonfiguration während der Kalibrierung

D1
D1
Y152201
D2
D2
2329-601-03
D3
D3
M2,U1,O1,L1,P5,B4,H1,NQ,E1,T1,S1,C2,A1,I1,J1,V1,X1,G0,K50,
D7
D7
NO DAC AVAILABLE